


EU Declaration of Conformity



Manufacturer: **Michell Instruments Limited**
48 Lancaster Way Business Park
Ely, Cambridgeshire
CB6 3NW. UK.



On behalf of the above named company, I declare that under our sole responsibility, on the date that the equipment accompanied by this declaration is placed on the market, it conforms with all technical and regulatory requirements of the listed EU Directives listed below

Easidew Comms Kit

2014/30/EU

EMC Directive

**2011/65/EU-2015/863/EU
Directive (RoHS3)**

Restriction of Hazardous Substances

EU Directive 2015/863/EU, Annex II states:

The restriction of DEHP, BBP, DBP and DIBP shall apply to.....monitoring and control instruments, including industrial monitoring and control instruments, from 22 July 2021.

and has been designed to be in conformance with the relevant sections of the following standards or other normative documents.

EN61326-1:2013

Electrical equipment for measurement, control and laboratory use –
EMC requirements –Class B (emissions) and Industrial Locations
(immunity).

EN61010-1:2010

Safety Requirements for Electrical Equipment for
Measurement, Control, and Laboratory Use - Part 1:
General Requirements

Andrew M.V. Stokes, Technical Director
Ely, UK
July 2019

Manufacturer: **Michell Instruments Limited**
48 Lancaster Way Business Park
Ely, Cambridgeshire
CB6 3NW. UK.



On behalf of the above named company, we the manufacturer declare that under our sole responsibility, on the date that the equipment accompanied by this declaration is placed on the market, the equipment conforms with all technical and regulatory requirements of the relevant sections of standards and/or other normative documents of the listed UKCA Regulations.

Easidew Comms Kit

The Electromagnetic Compatibility Regulations 2016.
UK Statutory Instrument (S.I. 2016 No.1091).

Is in conformity with the following Standard(s) or Normative Document(s):
EN61326-1:2013 *Electrical equipment for measurement, control and laboratory use - EMC requirements. Class B (emissions) and Industrial Locations (immunity).*

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012.

UK Statutory Instrument (S.I. 2012 No.3032).

States: *The restriction of DEHP, BBP, DBP and DIBP shall apply to,.....monitoring and control instruments, including industrial monitoring and control instruments, from 22 July 2021.*

The Electrical Equipment Safety Regulation 2016.

UK Statutory Instrument (S.I. 2016 1101).

BS/EN 61010-1:2010/A1:2019 Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Andrew M.V. Stokes, Technical Director
Ely, UK
July 2021